

1 2 1 '99 SEP -7 A10:03

Zymed Inc. 20 North Aviador Street Camarillo, California 93010-8348 800,235.5941 - 805.987.9611 Fax 805.987.9532

September 3, 1999

Dockets Management Branch (HFA09305) Food and Drug Administration 5630 Fishers Lane, rm. 1061 Rockville, MD 20852

Re: Comments for Docket No. 99N092075

## Comment 1

Reference: Section 1.1 An event has occurred

The term event is confusing because it generally refers to a device malfunction, serious injury, or death in the United States. Secondly, a device malfunction and inadequate design are not the same type of event. For example, an inadequate design parameter could cause a device malfunction, which may or may not be a reportable adverse event.

## Comment 2

1

Reference: Section 3 User Error

The term user error needs to be clearly defined. User error is a very subjective term. Its interpretation will vary based on the reporter of the adverse event. Users and manufacturers tend to define what is a user error differently from one another.

Gretel Lumley
QA Engineer

Zymed

20 N. Aviador

Camarillo, CA 93010

Place Labels in This Space 1997 1997 1997 1997 1997

Options: ALUPS Tracking Label and your address label

Place the Tracking Label to the left of the address label. Stamp your UPS Shipper Number or place parcel register tape above the address label.

**B** UPS Waybill

ZYMED MEDICAL INSTRUMENTATION (805) 987-9616 20 NORTH AVIADOR STREET CAMARILLO CA 93010

SHIP TO:

DOCKETS MANAGEMENT BRANCH #HFA09305 FOOD & DRUG ADMINISTRATION RM. 1061 5630 FISHERS LN. ROCKVILLE MD 20852



MD 2070-10



UPS NEXT DAY AIR SAFER TRACKING #: 1Z 867 429 13 4074 83

